

# **DIAL CALIBRATION TESTER (DCT-100)**

# DESCRIPTION


CNC dial calibration tester is ideal for high volume calibration of inductive and incremental probes and long length, dial indicators, comparators, lever type dials, plunger type dials, digital dial indicators & bore gauges. It saves the inspection time by 80% as compared to manual dial calibration tester. CNC DCT is software controlled user-friendly machine having high resolution measuring system along with motorized drive for fast calibration.

# **Features**

- Machine complies with Abbe's principle thus ensuring high accuracy
- Motorized measuring spindle drive for analog dial indicators with automation pre positioning Electronic handwheel for manual positoning of measuring spindle.
- Fully automatic calibration cycle for digital dials
- CNC-100 DCT can also be used horizontally for bore gauge (StickType) measurement
- Rigid and ergonomic design
- High accuracy optcal scale encoder system with error compensaton facility.
- User friendly sofware for easy operatons and Capable of maintaining Gauge History of Calibraton.

## Software : CNC DCT 100 V 1.0.0.

- For Lever Type Dial As Per IS 11498 : 1985
- For Plunger Type Dial As Per IS 2092 : 1983



Measurement Screen for Clockwise Direction of Plunger Dial Gauge



## **Technical Specification**

Measuring Range (mm)	100
Resolution mm	0.0001
Accuracy (at 20°C) (µm)	± (0.2 + L/200)
Drive Method	Electronic Motor
Measuring Unit	Numeric Jena Metallic Scale
Measurement	Semi Automatic / Fully
Speed	Max. 2 mm / s
Method	Automatic* / Fine Electonic Knob
Dimension	235 x 216 x 480 mm
Operating emperature (20°C)	20°C ± 0.5
Power Supply	110/230V/9V AC, 18 VA

### Accessories

- Standard accessory for lever type dial, microkators with 28 mm dia stem and bore gauge calibration
- · Holders for lever dial, digital dial and probes
- Analog and digital dial gauges.
- Faster by newest technology easier to handle.

ctivity No.: 106		Re	ading	S		
Temp. Correction					0.01	mm
C Required		A	C			
		B	м			0.01
Not Required						(
elect Combination Of Divisions	of Dial : PG-Temp	· •		Forward		C Reverse
<u>C</u> al Diagram	Clockwise	AntiClockwise		<u>R</u> epeatability	Per	formance
Accuracy	-	issible Error For L.C. 0.002 mm	Observed	Error - Clockwise (mm)	Observed Error - An	- ti-Clockwise (m
Accuracy Error In Any 1/10th Revolution.	-	issible Error For L.C. 0.002 mm 0.0010	Observed	Error - Clockwise (mm) 0.0010	Observed Error - An 0.00	ni-Clockwise (m 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions.	-	issible Error For L.C. 0.002 mm 0.0010 0.0040	Observed	Error - Clockwise (mm) 0.0010 0.0015	Observed Error - An 0.00 0.00	ti-Clockwise (m 15 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions. Error In Any 2 Revolutions.	-	issible Error For L.C. 0.002 mm 0.0010 0.0040 0.0060	Observed	Error - Clockwise (mm) 0.0010 0.0015 0.0015	Observed Error - An 0.00 0.00 0.00	ti-Clockwise (m 15 15 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions. Error In Any 2 Revolutions. Error In First 2 Revolutions.	Perm	issible Error For L.C. 0.002 mm 0.0010 0.0040 0.0060 NA	Observed	Error - Clockwise (mm) 0.0010 0.0015 0.0015 0.0015	0bserved Error - An 0.00 0.00 0.00 0.00	ni-Clockwise (m 15 15 15 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions. Error In Any 2 Revolutions. Error In First 2 Revolutions. Error In Any Larger Interval up To 1	Perm	issible Error For L.C. 0.002 mm 0.0010 0.0040 0.0060 NA 0.0180	Observed	Error - Clockwise (mm) 0.0010 0.0015 0.0015 0.0015 0.0015 0.0015	0bserved Error - An 0.00 0.00 0.00 0.00 0.00 0.00	ti-Clockwise (m 15 15 15 15 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions. Error In Any 2 Revolutions. Error In First 2 Revolutions.	Perm	issible Error For L.C. 0.002 mm 0.0010 0.0040 0.0060 NA	Observed	Error - Clockwise (mm) 0.0010 0.0015 0.0015 0.0015	0bserved Error - An 0.00 0.00 0.00 0.00	ti-Clockwise (m 15 15 15 15 15 15 15
Accuracy Error In Any 1/10th Revolution. Error In Any 1 Revolutions. Error In Any 2 Revolutions. Error In Any Larger Interval up To 1 Repeatability	Perm	issible Error For L.C. 0.002 mm 0.0010 0.0040 0.0060 NA 0.0180 0.0005	Observed	Error - Clockwise (mm) 0.0010 0.0015 0.0015 0.0015 0.0015 0.002 0.0010 0.0015	0bserved Error - An 0.00 0.00 0.00 0.00 0.00 0.00 0.00	4i-Clockwise (m 15 15 15 15 15 15 10 15

Parameter calculation as per IS: 2092-1983 of Measured Data for Plunger Dial Gauge

# **DIAL CALIBRATION TESTER**



### Features

- Precise and sensitve micrometer drum with non rotating anvil Robust construction with virtual stand for easy operation and clear visibility.
- Replaceable probe No need of dedicated probe as a master Design based on Abbe's principle
- Direct comparative measurement for required accuracy
- Horizontally it accommodates the Lever Type Dials & Dial bore gauges (Stick Type) for calibration
- Used in two different combinations with Digital Dial Gauge with 0.001mm resolution for calibration of 0.01mm least count dial gauges.
- Digital Display Unit and Probe with 0.0001mm resolution for 0.001 and 0.002mm least count dial gauges.
- Inch resolutions dial gauges can also be checked
- RS 232 Port for interface with computer and Software for recording and evaluation
- The DG Cal software provides calibration results. Accessories such as clamp, magnifying lens and holders are provided.

### **Technical Specification**

Features	Model DCT1	Model DCT2
Measuring System	Digital Display Unit & 10 mm to 25 mm Stroke Probe	Digital Dial
Calibration Ran e	0-25mm / 1"	0-25 mm / 1"
Repeatability	0.0003mm	0.001 mm
Displacement Error Over Full Range	0.0025mm	0.005 mm
Resolution	0.001/0.0001mm	0.001 mm



# Accurate Metrology And Automation Company

(Formerly "Aditya Engineering Company")

- Head Office : 67, Hadapsar Industrial Estate, Pune-411 013, Maharashtra, India.
  +91-20- 6606 9595, +91 86000 23615
- 💡 Works : A/P Khalad, Saswad Jejuri Road, Tal. Purandhar, Pune-412 301, India.
- +91 99700 02528 adityaengg@adityaengg.com GST: 27AACFA5688G1ZA +91 86000 21671 ● www.adityaengg.com PAN: AACFA5688G

 Calibration flats
 Calibration Results

 Results
 Calibration Results

 Prove the interval
 Note Matching

 Note Matching
 Note Matching

 The solution of the interval
 Note Matching

 Note Matching
 Note Matching

 Note Matching
 Note Matching

 Note Matching
 Note Matching

 Note of the interval
 Note Matching

 Note of the interval
 Note Network
 Note Network

 Note Network
 Note Network
 Note Network

 N

### **Features of DG Cal Software**

- Direct data acquisitionto computer
- Windows based and user friendly softare
- Graphical presentaton of Dial Indicator errors in forward and backward directions.
- Software is capable of maintaining Gauge history
- Plunger type dial as per IS 2092 : 1983
- Lever type dial as per IS 11498 : 1985
- Hysteresis result

## Application

- Calibraton of Plunger / Lever Type Dial of shank dia 8 mm
- Calibraton of Dial Bore Gauge (single point) up to 150 mm on std. accessories
- By using optonal accessories bore gauge can be checked up to 450 mm
- Calibration of microkators



Calibration of dial boe gauges

**Calibration of Microkators** 



Note : Due to continuous Improvements at our end, Specifications are subject to change without prior notice.